

Epithalon HPLC Analysis Report

Sample: **Ala-Glu-Asp-Gly**

Sample ID: **Epithalon**

Buffer A: 0.1% TFA in 100% water (v/v)

Buffer B: 0.1% TFA in 80% acetonitrile + 20% water (v/v)


Gradient: 100% Buffer A in 20min

Flow: 1ml/min

Wavelength: 220nm

Column: Venusil C18 5um 4.6*250mm

Chromatogram



Signal	Peak#	RT (min)	Area%	Area (mAU*s)	Height (mAU)
--------	-------	----------	-------	--------------	--------------

	1	11.643	99.69	5694272	123959
--	---	--------	-------	---------	--------

2	13.000	0.3062	17490	2175
<hr/>				
Total		100	5711762	126134

PROPERTY OF PEPTIDES WAREHOUSE